
Midmark M11 - v1.0.0
Total Cycles: 5077

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2024 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.4 F 0.0
2:00 147.3 F 0.0
4:00 157.7 F 0.5
6:00 188.5 F 0.7
8:00 206.4 F 2.8
10:00 219.4 F 6.4
12:00 229.2 F 10.4
14:00 233.2 F 8.3
16:00 238.8 F 11.3
18:00 235.9 F 9.0
20:00 240.4 F 10.3
22:00 247.4 F 14.0
24:00 259.0 F 20.2
26:00 269.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.8
2:00 271.6 F 28.3
3:00 272.5 F 28.8
4:00 271.4 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 208.9 F 0.0
10:00 222.5 F 0.0
15:00 235.4 F 0.0
20:00 229.5 F 0.0
25:00 231.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 27:03
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:05:37
CYCLE COMPLETE

11/11/2024 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____