
Midmark M11 - v1.0.0
Total Cycles: 3195

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2023 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.1 F 0.0
2:00 159.3 F 0.0
4:00 180.8 F 2.8
6:00 208.2 F 7.1
8:00 222.2 F 12.3
10:00 227.2 F 7.7
12:00 240.0 F 13.2
14:00 239.7 F 11.0
16:00 237.7 F 9.6
18:00 252.4 F 16.6
20:00 264.5 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.2 F 28.8
2:00 271.6 F 28.7
3:00 271.9 F 28.6
4:00 271.5 F 28.7

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 208.1 F 0.0
10:00 224.0 F 0.0
15:00 222.2 F 0.0
20:00 223.0 F 0.0
25:00 225.4 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

05/03/2023 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____