
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001085
10:04 12/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:46
STERILIZING 4:40
136.0 C 220kPa 4:40
136.4 C 225kPa 5:10
136.4 C 224kPa 5:40
136.5 C 224kPa 6:10
136.4 C 224kPa 6:40
136.3 C 224kPa 7:10
136.3 C 222kPa 7:40
136.4 C 222kPa 8:10
136.4 C 222kPa 8:10
VENTING 8:10
AIR DRYING 9:06
CYCLE COMPLETE 9:06

Digital Signature #
2724FE729FF6DF92
