
Midmark M11 - v1.0.5
Total Cycles: 2301

Name: RITTER2
Sterilizer ID: V2313228
04/25/2023 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 181.6 F 4.0
4:00 193.1 F 8.8
6:00 215.0 F 8.2
8:00 227.3 F 9.8
10:00 232.6 F 8.3
12:00 242.7 F 12.2
14:00 245.4 F 13.3
16:00 258.6 F 20.5
18:00 270.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.9 F 29.5
2:00 271.3 F 28.8
3:00 272.9 F 29.7
4:00 272.2 F 29.2

Min 271.2 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 216.1 F 0.0
10:00 238.2 F 0.0
15:00 231.7 F 0.0
20:00 229.7 F 0.0
25:00 226.5 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

04/25/2023 02:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____