
Midmark M11 - v1.0.0
Total Cycles: 3205

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2023 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.0 F 0.0
2:00 175.1 F 3.0
4:00 191.6 F 7.0
6:00 214.4 F 12.4
8:00 222.8 F 7.0
10:00 236.3 F 12.5
12:00 236.9 F 9.9
14:00 236.5 F 8.5
16:00 249.8 F 15.3
18:00 261.9 F 22.2
20:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.1
2:00 271.7 F 28.4
3:00 272.4 F 29.3
4:00 272.4 F 28.9

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 206.3 F 0.0
10:00 216.8 F 0.0
15:00 217.1 F 0.0
20:00 217.7 F 0.0
25:00 217.6 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

05/04/2023 04:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____