
Midmark M11 - v1.0.0
Total Cycles: 2204

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2022 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.1 F 0.0
2:00 75.7 F 0.0
4:00 79.4 F 0.0
6:00 107.8 F 0.5
8:00 139.6 F 2.3
10:00 165.0 F 5.3
12:00 186.4 F 8.5
14:00 197.3 F 12.4
16:00 213.9 F 4.5
18:00 224.3 F 8.2
20:00 233.7 F 12.1
22:00 233.2 F 8.1
24:00 244.5 F 13.4
26:00 242.4 F 11.4
28:00 254.1 F 17.4
30:00 264.5 F 23.6
32:00 271.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.0
2:00 271.5 F 28.1
3:00 272.8 F 29.0
4:00 271.7 F 28.2

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.4 F 0.0
5:00 207.2 F 0.0
10:00 221.5 F 0.0
15:00 218.9 F 0.0
20:00 224.3 F 0.0
25:00 225.2 F 0.0
30:00 227.0 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:05
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:01
CYCLE COMPLETE

08/08/2022 08:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____