
Midmark M11 - v1.0.5
Total Cycles: 2125

Name: RITTER2
Sterilizer ID: V2313228
03/08/2023 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.9 F 0.0
2:00 177.4 F 2.9
4:00 181.9 F 7.1
6:00 206.3 F 12.3
8:00 222.9 F 7.5
10:00 236.3 F 13.0
12:00 237.7 F 10.6
14:00 236.1 F 9.0
16:00 250.5 F 15.9
18:00 262.3 F 22.6
20:00 272.5 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.3
1:00 272.1 F 29.8
2:00 272.8 F 29.8
3:00 272.1 F 29.3
4:00 271.6 F 29.6

Min 271.4 F 29.1
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 236.5 F 0.0
10:00 230.1 F 0.0
15:00 215.8 F 0.0
20:00 209.5 F 0.0
25:00 213.5 F 0.0
30:00 210.8 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

03/08/2023 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____