
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000619
9:49 14/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:16
STERILIZING 4:25
136.4 C 224kPa 4:25
136.1 C 223kPa 4:55
136.1 C 222kPa 5:25
136.0 C 221kPa 5:55
136.2 C 221kPa 6:25
136.3 C 220kPa 6:55
136.3 C 220kPa 7:25
136.1 C 222kPa 7:55
136.1 C 222kPa 7:55
VENTING 7:55
AIR DRYING 8:55
CYCLE COMPLETE 9:00

Digital Signature #
510E8174A142EADA
