
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000040
12:45 19/02/2020
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 1:43
PRESSURIZING 3:39
STERILIZING 4:25
136.4 C 225kPa 4:25
136.2 C 224kPa 4:55
136.4 C 226kPa 5:25
136.2 C 224kPa 5:55
136.4 C 223kPa 6:25
136.4 C 222kPa 6:55
136.4 C 221kPa 7:25
136.4 C 227kPa 7:55
136.4 C 226kPa 8:25
136.4 C 226kPa 8:55
136.3 C 225kPa 9:25
136.4 C 221kPa 9:55
136.4 C 227kPa 10:25
136.1 C 223kPa 10:55
136.4 C 225kPa 11:25
136.4 C 223kPa 11:55
136.1 C 224kPa 12:25
136.3 C 225kPa 12:55
136.4 C 226kPa 13:25
136.5 C 227kPa 13:55
136.4 C 226kPa 14:25
136.4 C 226kPa 14:25
VENTING 14:25
AIR DRYING 15:10
CYCLE COMPLETE 30:10

Digital Signature #
286750D09215B80C
