
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001143
14:46 23/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:21
PRESSURIZING 4:22
STERILIZING 5:29
136.0 C 220kPa 5:29
136.4 C 226kPa 5:59
136.3 C 225kPa 6:29
136.3 C 225kPa 6:59
136.3 C 225kPa 7:29
136.3 C 224kPa 7:59
136.2 C 224kPa 8:29
136.4 C 226kPa 8:59
136.4 C 226kPa 8:59
VENTING 8:59
AIR DRYING 9:54
CYCLE COMPLETE 12:06

Digital Signature #
8C333FB6F4DBFE56
