
Midmark M11 - v1.0.5
Total Cycles: 2442

Name: RITTER2
Sterilizer ID: V2313228
05/31/2023 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.0 F 0.0
2:00 197.5 F 3.6
4:00 196.8 F 8.8
6:00 216.9 F 7.8
8:00 228.6 F 10.3
10:00 231.8 F 7.9
12:00 241.1 F 11.3
14:00 245.8 F 13.7
16:00 258.3 F 20.5
18:00 269.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.4 F 29.8
2:00 271.8 F 29.1
3:00 271.4 F 29.3
4:00 272.6 F 29.7

Min 271.2 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 214.9 F 0.0
10:00 212.5 F 0.0
15:00 225.4 F 0.0
20:00 222.7 F 0.0
25:00 220.7 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

05/31/2023 10:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____