
Midmark M11 - v1.0.0
Total Cycles: 2433

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2022 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 169.2 F 3.1
4:00 184.8 F 6.9
6:00 207.6 F 11.5
8:00 220.4 F 4.7
10:00 230.7 F 9.4
12:00 239.5 F 9.9
14:00 240.2 F 11.3
16:00 237.0 F 9.2
18:00 250.6 F 15.8
20:00 261.9 F 22.3
22:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.3 F 29.3
2:00 271.9 F 28.7
3:00 271.2 F 28.6
4:00 272.4 F 29.3

Min 271.1 F 28.4
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 226.1 F 0.0
10:00 230.1 F 0.0
15:00 229.0 F 0.0
20:00 227.9 F 0.0
25:00 228.3 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:28
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

10/14/2022 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____