
Midmark M11 - v1.0.0
Total Cycles: 3789

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2023 12:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.7 F 0.0
2:00 180.5 F 3.0
4:00 198.8 F 7.6
6:00 217.5 F 12.6
8:00 223.0 F 7.5
10:00 236.1 F 12.6
12:00 236.9 F 10.1
14:00 236.2 F 8.7
16:00 249.5 F 15.1
18:00 260.8 F 21.6
20:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.5 F 29.4
2:00 271.7 F 28.9
3:00 272.5 F 29.1
4:00 272.8 F 29.5

Min 270.9 F 28.3
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 243.0 F 0.0
10:00 231.3 F 0.0
15:00 224.4 F 0.0
20:00 197.7 F 0.0
25:00 194.6 F 0.0
30:00 238.9 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

09/18/2023 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____