
Midmark M11 - v1.0.0
Total Cycles: 2371

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2022 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.3 F 0.0
2:00 177.5 F 4.1
4:00 186.8 F 8.6
6:00 206.7 F 11.6
8:00 224.4 F 8.4
10:00 237.6 F 13.0
12:00 238.0 F 10.8
14:00 235.2 F 8.8
16:00 249.6 F 15.6
18:00 261.6 F 22.3
20:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.4 F 29.6
2:00 271.8 F 29.0
3:00 271.5 F 29.2
4:00 272.4 F 29.6

Min 271.2 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 223.1 F 0.0
10:00 233.1 F 0.0
15:00 236.8 F 0.0
20:00 232.4 F 0.0
25:00 226.0 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

09/26/2022 03:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____