
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.7uS / 1.7ppm
CYCLE NUMBER 000043
7:37 24/08/2020
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 10:39
STERILIZING 12:02
136.4 C 227kPa 12:02
135.8 C 220kPa 12:32
136.5 C 225kPa 13:02
136.3 C 225kPa 13:32
136.5 C 226kPa 14:02
136.4 C 225kPa 14:32
136.1 C 224kPa 15:02
136.3 C 226kPa 15:32
136.3 C 226kPa 16:02
136.2 C 224kPa 16:32
136.5 C 226kPa 17:02
136.3 C 226kPa 17:32
136.3 C 226kPa 18:02
VENTING 18:02
AIR DRYING 18:54
CYCLE COMPLETE 48:54

Digital Signature #
DA12DD45D434FF76
