

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001517  
10:18 19/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:10  
PRESSURIZING 3:14  
STERILIZING 4:19  
136.4 C 224kPa 4:19  
135.8 C 222kPa 4:49  
136.4 C 225kPa 5:19  
136.4 C 225kPa 5:49  
136.2 C 223kPa 6:19  
136.3 C 223kPa 6:49  
136.4 C 223kPa 7:19  
136.0 C 221kPa 7:49  
136.0 C 221kPa 7:49  
VENTING 7:49  
AIR DRYING 8:41  
CYCLE COMPLETE 8:54  
-----

Digital Signature #  
CD71DBA050D7A736  
-----