

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000565  
10:31 06/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:48  
STERILIZING 3:56  
136.0 C 228kPa 3:56  
136.4 C 233kPa 4:26  
136.5 C 231kPa 4:56  
136.2 C 231kPa 5:26  
136.4 C 232kPa 5:56  
136.4 C 231kPa 6:26  
136.3 C 231kPa 6:56  
136.1 C 231kPa 7:26  
136.1 C 231kPa 7:26  
VENTING 7:26  
AIR DRYING 8:16  
CYCLE COMPLETE 23:16  
-----

Digital Signature #  
F815540F4BAA6C28  
-----