
Midmark M11 - v1.0.0
Total Cycles: 997

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2021 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.1 F 0.0
2:00 208.3 F 3.4
4:00 215.1 F 8.6
6:00 223.6 F 9.8
8:00 233.3 F 9.3
10:00 240.0 F 9.6
12:00 243.6 F 12.7
14:00 242.2 F 11.3
16:00 255.1 F 17.9
18:00 266.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.6
2:00 271.7 F 28.6
3:00 272.1 F 28.5
4:00 271.6 F 28.5

Min 271.4 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.4 F 0.0
5:00 227.1 F 0.0
10:00 225.7 F 0.0
15:00 226.9 F 0.0
20:00 241.8 F 0.0
25:00 231.1 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

07/12/2021 02:47 PM
ACCEPTED: X
REJECTED:
Operator: U01