
Midmark M11 - v1.0.0
Total Cycles: 4803

Name: RITTER 1
Sterilizer ID: V2178198
08/12/2024 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.3 F 0.0
2:00 166.1 F 1.8
4:00 183.2 F 5.2
6:00 200.3 F 9.5
8:00 214.6 F 8.5
10:00 223.1 F 8.3
12:00 235.6 F 13.3
14:00 237.4 F 10.1
16:00 236.1 F 8.4
18:00 248.1 F 14.2
20:00 259.2 F 20.3
22:00 269.5 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.9 F 28.9
2:00 272.3 F 29.2
3:00 271.9 F 28.7
4:00 272.5 F 29.2

Min 271.4 F 28.5
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 236.3 F 0.0
10:00 227.7 F 0.0
15:00 226.3 F 0.0
20:00 197.5 F 0.0
25:00 180.8 F 0.0
30:00 190.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 23:49
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:02:27
CYCLE COMPLETE

08/12/2024 12:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____