

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001803  
9:04 26/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:09  
STERILIZING 3:11  
136.0 C 220kPa 3:11  
136.3 C 225kPa 3:41  
136.2 C 223kPa 4:11  
136.1 C 222kPa 4:41  
136.3 C 224kPa 5:11  
136.4 C 225kPa 5:41  
136.4 C 225kPa 6:11  
136.4 C 225kPa 6:41  
136.4 C 225kPa 6:41  
VENTING 6:41  
AIR DRYING 7:32  
CYCLE COMPLETE 37:32

-----  
Digital Signature #  
B09D77E2685A92AB  
-----