

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5295

Name: RITTER 1  
Sterilizer ID: V2178198  
01/29/2025 05:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.7 F 0.0  
2:00 149.4 F 2.5  
4:00 169.7 F 6.2  
6:00 190.4 F 10.2  
8:00 212.0 F 5.1  
10:00 218.1 F 6.5  
12:00 230.7 F 10.5  
14:00 231.7 F 8.2  
16:00 241.0 F 11.5  
18:00 237.0 F 9.2  
20:00 250.0 F 15.3  
22:00 260.7 F 21.4  
24:00 270.6 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.5 F 29.5  
2:00 271.6 F 28.8  
3:00 272.3 F 29.6  
4:00 272.3 F 29.1

Min 271.3 F 28.7  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.1 F 0.0  
5:00 217.4 F 0.0  
10:00 231.0 F 0.0  
15:00 226.7 F 0.0  
20:00 222.4 F 0.0  
25:00 223.6 F 0.0  
30:00 240.7 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 25:36  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:04:23  
CYCLE COMPLETE

01/29/2025 06:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_