

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2673

Name: RITTER2  
Sterilizer ID: V2313228  
07/20/2023 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.0 F 0.0  
2:00 160.4 F 0.1  
4:00 160.1 F 3.1  
6:00 183.5 F 7.5  
8:00 206.4 F 12.5  
10:00 220.8 F 7.5  
12:00 234.4 F 12.7  
14:00 235.5 F 9.8  
16:00 236.0 F 8.9  
18:00 247.3 F 14.4  
20:00 259.0 F 21.0  
22:00 269.7 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 272.2 F 29.8  
2:00 271.3 F 29.0  
3:00 271.8 F 29.8  
4:00 272.0 F 29.6

Min 271.1 F 29.0  
Max 272.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 206.4 F 0.0  
10:00 218.3 F 0.0  
15:00 225.7 F 0.0  
20:00 222.9 F 0.0  
25:00 225.0 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:00  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:10  
CYCLE COMPLETE

07/20/2023 10:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_