

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2658

Name: RITTER 1  
Sterilizer ID: V2178198  
12/08/2022 10:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.2 F 0.0  
2:00 188.2 F 2.5  
4:00 206.2 F 6.2  
6:00 211.7 F 10.9  
8:00 221.4 F 5.9  
10:00 232.9 F 11.3  
12:00 234.0 F 8.2  
14:00 241.8 F 11.0  
16:00 245.9 F 13.0  
18:00 258.3 F 19.7  
20:00 269.5 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 28.9  
2:00 271.4 F 28.2  
3:00 272.6 F 29.1  
4:00 271.8 F 28.4  
  
Min 271.4 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.2 F 0.0  
5:00 228.1 F 0.0  
10:00 237.2 F 0.0  
15:00 229.4 F 0.0  
20:00 226.6 F 0.0  
25:00 228.4 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 21:01  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

12/08/2022 11:09 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_