
Midmark M11 - v1.0.5
Total Cycles: 2581

Name: RITTER2
Sterilizer ID: V2313228
06/28/2023 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.9 F 0.0
2:00 199.8 F 4.1
4:00 207.6 F 9.9
6:00 223.2 F 6.7
8:00 235.1 F 11.7
10:00 236.3 F 9.9
12:00 236.2 F 8.8
14:00 250.9 F 16.3
16:00 263.4 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.1
1:00 272.4 F 29.9
2:00 272.2 F 29.2
3:00 271.9 F 29.6
4:00 272.6 F 29.7

Min 271.5 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 217.5 F 0.0
10:00 227.9 F 0.0
15:00 229.4 F 0.0
20:00 225.5 F 0.0
25:00 224.2 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 17:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:12
CYCLE COMPLETE

06/28/2023 01:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____