
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000965
17:19 07/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:53
STERILIZING 5:26
136.5 C 232kPa 5:26
136.3 C 232kPa 5:56
136.4 C 233kPa 6:26
136.2 C 231kPa 6:56
136.4 C 233kPa 7:26
136.3 C 232kPa 7:56
136.3 C 232kPa 8:26
136.4 C 230kPa 8:56
136.4 C 230kPa 8:56
VENTING 8:56
AIR DRYING 9:53
CYCLE COMPLETE 24:53

Digital Signature #
1A78DCC38B3E639E
