
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000050
9:14 06/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:41
STERILIZING 3:58
136.4 C 225kPa 3:58
136.4 C 224kPa 4:28
136.2 C 222kPa 4:58
136.4 C 221kPa 5:28
136.4 C 223kPa 5:58
136.4 C 223kPa 6:28
135.9 C 221kPa 6:58
136.3 C 223kPa 7:28
136.3 C 223kPa 7:28
VENTING 7:28
AIR DRYING 8:21
CYCLE COMPLETE 18:53

Digital Signature #
23F465C74F1AA5ED
