
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001573
12:38 08/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:30
STERILIZING 3:33
136.0 C 220kPa 3:33
136.4 C 225kPa 4:03
136.4 C 226kPa 4:33
136.4 C 224kPa 5:03
136.4 C 222kPa 5:33
136.1 C 224kPa 6:03
136.1 C 222kPa 6:33
136.4 C 225kPa 7:03
136.4 C 225kPa 7:03
VENTING 7:03
AIR DRYING 7:54
CYCLE COMPLETE 8:10

Digital Signature #
6D8471C0119C9A3D
