
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000701
11:57 16/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:32
PRESSURIZING 1:40
STERILIZING 2:36
136.0 C 227kPa 2:36
136.4 C 233kPa 3:06
136.1 C 230kPa 3:36
136.4 C 232kPa 4:06
136.4 C 232kPa 4:36
136.1 C 230kPa 5:06
136.4 C 233kPa 5:36
136.2 C 231kPa 6:06
136.2 C 231kPa 6:06
VENTING 6:06
AIR DRYING 6:58
CYCLE COMPLETE 7:00

Digital Signature #
0C2E3F799393C36C
