
Midmark M11 - v1.0.0
Total Cycles: 4896

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2024 06:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 187.0 F 1.7
4:00 214.7 F 7.4
6:00 228.9 F 13.2
8:00 233.4 F 9.6
10:00 236.6 F 8.0
12:00 243.4 F 11.6
14:00 246.4 F 13.3
16:00 260.4 F 20.9
18:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.9 F 28.0
2:00 272.4 F 28.4
3:00 272.7 F 28.8
4:00 272.0 F 28.5

Min 271.6 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 233.8 F 0.0
10:00 235.4 F 0.0
15:00 227.2 F 0.0
20:00 223.3 F 0.0
25:00 224.9 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:14
CYCLE COMPLETE

09/11/2024 07:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____