
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000676
14:11 14/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:13
STERILIZING 3:16
136.4 C 230kPa 3:16
136.2 C 231kPa 3:47
136.4 C 230kPa 4:17
136.2 C 231kPa 4:47
136.4 C 230kPa 5:17
136.3 C 233kPa 5:47
136.4 C 230kPa 6:17
136.4 C 232kPa 6:47
136.4 C 232kPa 6:47
VENTING 6:47
AIR DRYING 7:39
CYCLE COMPLETE 7:45

Digital Signature #
B4D304054B1EBF83
