
Midmark M11 - v1.0.0
Total Cycles: 917

Name: RITTER 1
Sterilizer ID: V2178198
06/18/2021 01:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.4 F 0.0
2:00 112.9 F 0.0
4:00 136.2 F 0.1
6:00 156.9 F 1.1
8:00 180.9 F 3.5
10:00 201.2 F 6.4
12:00 208.6 F 9.9
14:00 221.3 F 10.1
16:00 232.8 F 8.8
18:00 243.0 F 13.4
20:00 239.5 F 10.6
22:00 236.1 F 8.5
24:00 249.2 F 14.7
26:00 260.0 F 20.8
28:00 270.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.5 F 28.9
2:00 272.4 F 29.3
3:00 271.8 F 28.6
4:00 272.6 F 29.2

Min 271.6 F 28.6
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.5 F 0.0
5:00 235.8 F 0.0
10:00 215.1 F 0.0
15:00 207.3 F 0.0
20:00 216.0 F 0.0
25:00 211.8 F 0.0
30:00 211.0 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 29:44
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:08:13
CYCLE COMPLETE

06/18/2021 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____