

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3017

Name: RITTER 1  
Sterilizer ID: V2178198  
03/20/2023 08:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.8 F 0.0  
2:00 175.5 F 2.7  
4:00 185.1 F 6.3  
6:00 208.5 F 10.9  
8:00 219.7 F 5.6  
10:00 234.9 F 10.9  
12:00 232.8 F 7.5  
14:00 244.4 F 12.7  
16:00 243.9 F 12.3  
18:00 256.6 F 19.0  
20:00 267.9 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.1  
2:00 271.4 F 28.4  
3:00 272.6 F 29.3  
4:00 272.0 F 28.7  
  
Min 271.4 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 229.5 F 0.0  
10:00 231.1 F 0.0  
15:00 221.5 F 0.0  
20:00 219.4 F 0.0  
25:00 220.4 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:38  
CYCLE COMPLETE

03/20/2023 09:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_