
Midmark M11 - v1.0.5
Total Cycles: 3580

Name: RITTER2
Sterilizer ID: V2313228
05/02/2024 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.9 F 0.0
2:00 77.5 F 0.0
4:00 86.7 F 0.4
6:00 113.5 F 2.2
8:00 152.6 F 5.0
10:00 190.6 F 8.1
12:00 193.4 F 12.1
14:00 209.2 F 5.1
16:00 222.1 F 8.8
18:00 233.0 F 12.9
20:00 231.9 F 8.8
22:00 241.8 F 12.0
24:00 241.9 F 12.0
26:00 253.4 F 17.9
28:00 263.7 F 24.0
30:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.2 F 30.2
2:00 271.8 F 29.7
3:00 272.6 F 29.9
4:00 272.8 F 30.4

Min 270.9 F 29.1
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 202.0 F 0.0
10:00 189.1 F 0.0
15:00 196.2 F 0.0
20:00 219.2 F 0.0
25:00 222.4 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:10
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:08:26
CYCLE COMPLETE

05/02/2024 09:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____