
Midmark M11 - v1.0.5
Total Cycles: 2151

Name: RITTER2
Sterilizer ID: V2313228
03/15/2023 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 175.7 F 4.1
4:00 186.3 F 8.6
6:00 211.5 F 8.7
8:00 224.9 F 8.9
10:00 234.2 F 10.2
12:00 241.1 F 12.5
14:00 241.2 F 11.5
16:00 254.7 F 18.4
18:00 266.5 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.8 F 29.8
2:00 271.6 F 29.0
3:00 272.2 F 29.9
4:00 272.4 F 29.6

Min 271.3 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 219.4 F 0.0
10:00 207.8 F 0.0
15:00 207.7 F 0.0
20:00 207.5 F 0.0
25:00 211.6 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

03/15/2023 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____