
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001298
10:31 14/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:28
PRESSURIZING 4:29
STERILIZING 6:12
136.5 C 225kPa 6:12
136.4 C 224kPa 6:42
136.3 C 222kPa 7:12
136.2 C 223kPa 7:42
136.3 C 224kPa 8:12
136.4 C 224kPa 8:42
136.4 C 224kPa 9:12
136.4 C 222kPa 9:42
136.4 C 222kPa 9:42
VENTING 9:42
AIR DRYING 10:41
CYCLE COMPLETE 25:41

Digital Signature #
B250A273B4CAFBBB
