
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004347
13:02 13/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:47
STERILIZING 4:18
136.4 C 221kPa 4:18
136.3 C 222kPa 4:48
136.4 C 224kPa 5:18
136.2 C 224kPa 5:48
136.4 C 224kPa 6:18
135.8 C 221kPa 6:48
136.4 C 221kPa 7:18
136.4 C 224kPa 7:48
136.4 C 224kPa 7:48
VENTING 7:48
AIR DRYING 8:40
CYCLE COMPLETE 8:45

Digital Signature #
DD5D3107B2A47516
