

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3538

Name: RITTER2  
Sterilizer ID: V2313228  
04/17/2024 03:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.4 F 0.0  
2:00 118.1 F 0.3  
4:00 142.9 F 2.7  
6:00 165.0 F 5.7  
8:00 188.0 F 9.0  
10:00 205.0 F 12.8  
12:00 216.4 F 6.7  
14:00 226.3 F 11.2  
16:00 228.9 F 7.7  
18:00 242.5 F 13.3  
20:00 241.4 F 11.9  
22:00 253.4 F 18.0  
24:00 264.2 F 24.5  
26:00 271.8 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.9  
1:00 272.8 F 30.3  
2:00 271.7 F 30.0  
3:00 272.7 F 30.3  
4:00 271.6 F 29.8

Min 271.6 F 29.7  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 219.8 F 0.0  
10:00 205.9 F 0.0  
15:00 198.3 F 0.0  
20:00 196.5 F 0.0  
25:00 198.4 F 0.0  
30:00 208.4 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 26:58  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:05:40  
CYCLE COMPLETE

04/17/2024 04:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_