
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007671
11:05 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:50
STERILIZING 3:49
136.0 C 220kPa 3:49
136.0 C 221kPa 4:19
136.2 C 218kPa 4:49
136.4 C 224kPa 5:19
136.4 C 222kPa 5:49
136.3 C 223kPa 6:19
136.4 C 222kPa 6:49
136.3 C 219kPa 7:19
136.3 C 219kPa 7:19
VENTING 7:19
AIR DRYING 8:10
CYCLE COMPLETE 8:40

Digital Signature #
6450314A5CEB6839
