
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000597
11:50 07/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:36
STERILIZING 4:41
136.3 C 222kPa 4:41
136.2 C 220kPa 5:11
136.3 C 222kPa 5:41
136.4 C 221kPa 6:11
136.4 C 221kPa 6:41
136.2 C 221kPa 7:11
136.4 C 225kPa 7:41
136.3 C 224kPa 8:11
136.3 C 224kPa 8:11
VENTING 8:11
AIR DRYING 9:05
CYCLE COMPLETE 9:19

Digital Signature #
9A9FA729FE5E1467
