
Midmark M11 - v1.0.5
Total Cycles: 1988

Name: RITTER2
Sterilizer ID: V2313228
01/31/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 169.8 F 2.7
4:00 176.3 F 7.1
6:00 198.0 F 12.3
8:00 224.9 F 7.4
10:00 238.0 F 12.8
12:00 238.4 F 10.1
14:00 237.3 F 8.6
16:00 250.4 F 15.4
18:00 262.0 F 22.3
20:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 273.0 F 30.0
2:00 272.1 F 29.5
3:00 271.1 F 29.4
4:00 272.9 F 29.8

Min 271.0 F 29.1
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 226.6 F 0.0
10:00 233.9 F 0.0
15:00 230.3 F 0.0
20:00 227.9 F 0.0
25:00 230.7 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

01/31/2023 03:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____