

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3621

Name: RITTER2  
Sterilizer ID: V2313228  
05/15/2024 01:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.1 F 0.0  
2:00 116.0 F 1.1  
4:00 139.4 F 3.5  
6:00 161.6 F 6.5  
8:00 182.6 F 10.2  
10:00 205.3 F 7.1  
12:00 216.4 F 6.9  
14:00 228.0 F 11.1  
16:00 227.6 F 7.0  
18:00 239.4 F 12.5  
20:00 238.9 F 10.7  
22:00 250.5 F 16.4  
24:00 261.0 F 22.4  
26:00 270.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.6 F 30.2  
2:00 271.9 F 29.6  
3:00 271.0 F 29.5  
4:00 272.4 F 30.5

Min 271.0 F 29.3  
Max 272.9 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 221.6 F 0.0  
10:00 197.4 F 0.0  
15:00 193.2 F 0.0  
20:00 191.5 F 0.0  
25:00 203.3 F 0.0  
30:00 209.5 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 26:53  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:05:37  
CYCLE COMPLETE

05/15/2024 02:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_