
Midmark M11 - v1.0.5
Total Cycles: 829

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 183.9 F 3.9
4:00 199.3 F 9.4
6:00 219.9 F 7.1
8:00 232.8 F 10.9
10:00 234.0 F 8.9
12:00 238.6 F 9.8
14:00 249.5 F 15.4
16:00 262.2 F 22.6
18:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.4
2:00 271.5 F 29.0
3:00 272.7 F 29.5
4:00 271.6 F 28.8

Min 271.5 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 222.4 F 0.0
10:00 223.7 F 0.0
15:00 226.7 F 0.0
20:00 229.7 F 0.0
25:00 227.3 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 2:21
HEATING: 18:16
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:43
CYCLE COMPLETE

02/23/2022 03:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____