
Midmark M11 - v1.0.5
Total Cycles: 3509

Name: RITTER2
Sterilizer ID: V2313228
04/09/2024 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.3 F 0.0
2:00 86.4 F 0.0
4:00 91.8 F 0.1
6:00 135.6 F 0.3
8:00 173.0 F 2.4
10:00 190.5 F 5.4
12:00 206.2 F 8.6
14:00 223.7 F 12.2
16:00 224.6 F 6.5
18:00 234.0 F 10.8
20:00 235.7 F 8.0
22:00 242.3 F 13.0
24:00 241.7 F 11.8
26:00 253.6 F 17.8
28:00 264.1 F 24.1
30:00 271.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.9 F 29.8
2:00 271.6 F 29.1
3:00 271.8 F 29.7
4:00 272.9 F 29.6

Min 271.2 F 29.0
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 203.7 F 0.0
10:00 199.9 F 0.0
15:00 194.4 F 0.0
20:00 192.2 F 0.0
25:00 193.5 F 0.0
30:00 194.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:09
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

04/09/2024 03:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____