

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1389

Name: RITTER 1  
Sterilizer ID: V2178198  
11/23/2021 02:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	147.5 F	0.0
2:00	149.9 F	1.0
4:00	159.5 F	4.6
6:00	185.7 F	9.1
8:00	201.2 F	6.6
10:00	221.1 F	6.9
12:00	235.0 F	11.9
14:00	234.3 F	8.5
16:00	241.4 F	10.7
18:00	246.0 F	13.1

C383: HEATUP MODE  
DOOR OPEN  
CHAMBER TEMP 254F  
CHAMBER PRESS 17.1PSI  
\*\* ITEMS NOT STERILE \*\*

11/23/2021 02:37 PM