
Midmark M11 - v1.0.0
Total Cycles: 3321

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2023 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.9 F 0.0
2:00 200.9 F 4.5
4:00 214.6 F 10.2
6:00 227.4 F 6.2
8:00 238.1 F 11.8
10:00 237.3 F 9.7
12:00 238.6 F 8.6
14:00 250.3 F 15.4
16:00 262.5 F 22.3
18:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.2
3:00 272.6 F 29.1
4:00 272.0 F 28.5

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 224.5 F 0.0
10:00 228.1 F 0.0
15:00 225.8 F 0.0
20:00 226.6 F 0.0
25:00 226.2 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 2:29
HEATING: 18:16
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:60
CYCLE COMPLETE

06/01/2023 12:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____