
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000089
11:52 13/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:35
STERILIZING 3:36
136.4 C 221kPa 3:36
136.0 C 223kPa 4:06
136.5 C 224kPa 4:36
136.2 C 223kPa 5:06
136.4 C 221kPa 5:36
135.8 C 221kPa 6:06
136.1 C 222kPa 6:36
136.2 C 219kPa 7:06
136.2 C 219kPa 7:06
VENTING 7:06
AIR DRYING 7:52
CYCLE COMPLETE 7:53

Digital Signature #
EA99D50D5A03E1A2
