

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3460

Name: RITTER 1  
Sterilizer ID: V2178198  
06/28/2023 01:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.2 F 0.0  
2:00 189.2 F 3.9  
4:00 201.9 F 8.5  
6:00 223.4 F 8.7  
8:00 232.1 F 9.5  
10:00 235.2 F 8.9  
12:00 244.5 F 13.0  
14:00 243.1 F 11.8  
16:00 256.3 F 18.7  
18:00 268.0 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.4 F 28.8  
2:00 271.4 F 28.3  
3:00 272.6 F 28.9  
4:00 271.5 F 28.2  
  
Min 271.3 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 237.0 F 0.0  
10:00 217.3 F 0.0  
15:00 226.4 F 0.0  
20:00 210.5 F 0.0  
25:00 224.7 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 19:17  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:18  
CYCLE COMPLETE

06/28/2023 02:01 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_