
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000172
11:39 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:40
PRESSURIZING 4:40
STERILIZING 5:47
136.0 C 222kPa 5:47
136.1 C 223kPa 6:17
136.3 C 225kPa 6:47
136.4 C 225kPa 7:17
136.4 C 226kPa 7:47
136.4 C 225kPa 8:17
136.2 C 223kPa 8:47
136.4 C 225kPa 9:17
136.4 C 225kPa 9:17
VENTING 9:17
AIR DRYING 10:20
CYCLE COMPLETE 25:20

Digital Signature #
31D3D53EF22C9B73
