

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1919

Name: RITTER 1  
Sterilizer ID: V2178198  
05/09/2022 10:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.8 F 0.0  
2:00 146.9 F 0.2  
4:00 163.9 F 2.0  
6:00 189.8 F 6.2  
8:00 211.5 F 10.9  
10:00 226.1 F 5.1  
12:00 235.3 F 10.2  
14:00 235.6 F 8.0  
16:00 244.0 F 12.3  
18:00 241.0 F 10.5  
20:00 254.0 F 16.9  
22:00 265.0 F 23.5  
24:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.8 F 28.7  
2:00 271.7 F 28.6  
3:00 272.5 F 28.6  
4:00 271.7 F 28.6

Min 271.4 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
5:00 232.1 F 0.0  
10:00 216.5 F 0.0  
15:00 212.2 F 0.0  
20:00 197.1 F 0.0  
25:00 227.3 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:37  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:02:40  
CYCLE COMPLETE

05/09/2022 11:20 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_