
Midmark M11 - v1.0.5
Total Cycles: 3661

Name: RITTER2
Sterilizer ID: V2313228
06/03/2024 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 72.5 F 0.1
4:00 109.5 F 0.5
6:00 104.9 F 2.2
8:00 134.2 F 4.8
10:00 161.5 F 7.7
12:00 182.7 F 10.8
14:00 203.9 F 5.2
16:00 212.4 F 6.1
18:00 221.8 F 9.5
20:00 230.7 F 13.5
22:00 233.3 F 9.3
24:00 240.5 F 11.3
26:00 242.4 F 12.1
28:00 253.3 F 17.7
30:00 263.1 F 23.5
32:00 272.2 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.1
1:00 272.5 F 30.0
2:00 271.5 F 29.1
3:00 272.5 F 29.6
4:00 271.7 F 29.5

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 219.5 F 0.0
10:00 210.7 F 0.0
15:00 200.6 F 0.0
20:00 205.2 F 0.0
25:00 208.5 F 0.0
30:00 211.2 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:21
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:10:28
CYCLE COMPLETE

06/03/2024 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____